
Midmark M11 - v1.0.5
Total Cycles: 1469

Name: RITTER2
Sterilizer ID: V2313228
09/08/2022 10:35 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.9 F 0.0
2:00 194.0 F 3.7
4:00 196.8 F 9.4
6:00 214.5 F 6.9
8:00 230.6 F 11.1
10:00 233.0 F 8.7
12:00 239.2 F 10.1
14:00 246.3 F 14.3
16:00 258.9 F 21.2
18:00 270.1 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.5
1:00 272.9 F 30.1
2:00 271.9 F 29.3
3:00 272.8 F 30.3
4:00 272.5 F 29.8

Min 271.7 F 29.3
Max 273.0 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.4 F 0.0
5:00 203.6 F 0.0
10:00 213.5 F 0.0
15:00 219.5 F 0.0
20:00 223.5 F 0.0
25:00 226.6 F 0.0
30:00 228.3 F 0.0
Drying Complete

FILLING: 2:19
HEATING: 18:54
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:23
CYCLE COMPLETE

09/08/2022 11:33 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____