

-----  
Midmark M11 - v1.0.5  
Total Cycles: 520

Name: RITTER2  
Sterilizer ID: V2313228  
11/08/2021 09:53 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 191.3 F 0.0  
2:00 186.4 F 2.4  
4:00 188.0 F 7.3  
6:00 209.4 F 13.5  
8:00 220.5 F 9.7  
10:00 231.2 F 8.5  
12:00 239.9 F 11.5  
14:00 245.2 F 13.6  
16:00 258.0 F 20.5  
18:00 268.3 F 27.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 29.4  
1:00 272.0 F 30.0  
2:00 271.2 F 29.6  
3:00 272.0 F 29.9  
4:00 271.0 F 29.4  
  
Min 271.0 F 29.3  
Max 272.3 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.3 F 0.0  
5:00 236.1 F 0.0  
10:00 226.7 F 0.0  
15:00 228.3 F 0.0  
20:00 224.1 F 0.0  
25:00 229.5 F 0.0  
30:00 229.2 F 0.0  
Drying Complete

FILLING: 1:55  
HEATING: 19:03  
STERILIZING: 4:00  
VENTING: 3:00  
DRYING: 30:00  
TOTAL CYCLE: 00:57:58  
CYCLE COMPLETE

11/08/2021 10:53 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01