

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3482

Name: RITTER 1  
Sterilizer ID: V2178198  
07/04/2023 02:36 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 185.0 F 0.0  
2:00 185.3 F 3.7  
4:00 202.0 F 8.7  
6:00 219.5 F 9.1  
8:00 230.2 F 9.5  
10:00 235.3 F 8.9  
12:00 244.8 F 12.9  
14:00 243.3 F 11.8  
16:00 256.5 F 18.6  
18:00 268.0 F 25.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.5  
1:00 272.2 F 28.6  
2:00 272.7 F 29.1  
3:00 271.7 F 28.6  
4:00 272.1 F 28.8  
  
Min 271.6 F 28.4  
Max 272.8 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.8 F 0.0  
5:00 202.4 F 0.0  
10:00 233.2 F 0.0  
15:00 229.8 F 0.0  
20:00 226.6 F 0.0  
25:00 216.2 F 0.0  
30:00 212.6 F 0.0  
Drying Complete

FILLING: 1:60  
HEATING: 19:58  
STERILIZING: 4:00  
VENTING: 3:21  
DRYING: 30:00  
TOTAL CYCLE: 00:59:18  
CYCLE COMPLETE

07/04/2023 03:35 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_