
Midmark M11 - v1.0.5
Total Cycles: 832

Name: RITTER2
Sterilizer ID: V2313228
02/23/2022 05:20 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 167.8 F 0.0
2:00 171.9 F 3.3
4:00 197.7 F 7.8
6:00 214.6 F 12.7
8:00 224.8 F 7.5
10:00 237.2 F 12.6
12:00 238.3 F 10.7
14:00 237.6 F 9.6
16:00 251.6 F 16.4
18:00 262.8 F 23.0
20:00 272.7 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.8
1:00 272.6 F 29.5
2:00 271.5 F 28.7
3:00 272.5 F 29.6
4:00 272.2 F 29.1

Min 271.4 F 28.7
Max 272.9 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.2 F 0.0
5:00 217.2 F 0.0
10:00 208.4 F 0.0
15:00 203.6 F 0.0
20:00 199.6 F 0.0
25:00 195.8 F 0.0
30:00 192.2 F 0.0
Drying Complete

FILLING: 2:36
HEATING: 20:18
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:00:16
CYCLE COMPLETE

02/24/2022 07:38 AM
ACCEPTED: X
REJECTED:
Operator: U02