
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000215
9:51 22/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:59
PRESSURIZING 3:59
STERILIZING 5:31
136.4 C 233kPa 5:31
136.3 C 233kPa 6:01
136.4 C 232kPa 6:31
136.4 C 232kPa 7:01
136.3 C 233kPa 7:31
136.4 C 233kPa 8:01
136.4 C 233kPa 8:31
136.4 C 233kPa 9:01
136.4 C 233kPa 9:01
VENTING 9:01
AIR DRYING 10:03
CYCLE COMPLETE 14:51

Digital Signature #
8E104F349A312D4D
