

-----  
Midmark M11 - v1.0.0  
Total Cycles: 943

Name: RITTER 1  
Sterilizer ID: V2178198  
06/25/2021 01:13 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 93.0 F 0.0  
2:00 102.1 F 0.0  
4:00 111.0 F 0.0  
6:00 128.7 F 0.1  
8:00 155.8 F 1.3  
10:00 178.1 F 3.4  
12:00 191.6 F 5.4  
14:00 202.9 F 7.8  
16:00 214.0 F 10.8  
18:00 220.9 F 7.2  
20:00 229.7 F 7.8  
22:00 239.7 F 11.9  
24:00 237.4 F 9.4  
26:00 241.8 F 10.5  
28:00 246.9 F 13.4  
30:00 257.3 F 19.1  
32:00 266.9 F 24.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.1  
1:00 272.1 F 28.7  
2:00 272.5 F 28.7  
3:00 272.1 F 28.3  
4:00 271.5 F 28.0  
  
Min 271.4 F 28.0  
Max 272.6 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 178.5 F 0.0  
5:00 215.8 F 0.0  
10:00 220.5 F 0.0  
15:00 212.7 F 0.0  
20:00 209.8 F 0.0  
25:00 214.9 F 0.0  
30:00 218.0 F 0.0  
  
Drying Complete

FILLING: 0:57  
HEATING: 33:47  
STERILIZING: 4:00  
VENTING: 3:44  
DRYING: 30:00  
TOTAL CYCLE: 01:12:28  
CYCLE COMPLETE

06/25/2021 02:25 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_