
Midmark M11 - v1.0.5
Total Cycles: 2821

Name: RITTER2
Sterilizer ID: V2313228
08/31/2023 08:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.1 F 0.0
2:00 70.4 F 0.0
4:00 76.6 F 0.0
6:00 108.4 F 0.0
8:00 143.9 F 0.2
10:00 165.9 F 2.1
12:00 189.6 F 4.5
14:00 221.6 F 7.7
16:00 231.1 F 11.3
18:00 222.2 F 5.7
20:00 236.0 F 10.4
22:00 233.8 F 8.5
24:00 243.5 F 12.7
26:00 241.6 F 11.2
28:00 253.6 F 17.3
30:00 264.1 F 23.6
32:00 271.7 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.6
1:00 272.4 F 28.9
2:00 272.4 F 29.3
3:00 271.3 F 28.6
4:00 272.4 F 29.0

Min 271.2 F 28.4
Max 272.6 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.0 F 0.0
5:00 231.3 F 0.0
10:00 230.5 F 0.0
15:00 182.1 F 0.0
20:00 178.5 F 0.0
25:00 185.7 F 0.0
30:00 197.8 F 0.0

Drying Complete

FILLING: 0:50
HEATING: 32:06
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:10:10
CYCLE COMPLETE

08/31/2023 10:01 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____