

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1708

Name: RITTER2  
Sterilizer ID: V2313228  
11/10/2022 01:12 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 161.1 F 0.0  
2:00 159.3 F 0.2  
4:00 190.7 F 0.9  
6:00 213.1 F 5.5  
8:00 227.4 F 10.4  
10:00 231.6 F 7.0  
12:00 240.2 F 12.8  
14:00 240.1 F 11.2  
16:00 239.7 F 10.5  
18:00 254.1 F 17.8  
20:00 266.2 F 25.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.9  
1:00 272.4 F 29.5  
2:00 271.4 F 28.7  
3:00 272.4 F 29.7  
4:00 272.0 F 29.2  
  
Min 271.2 F 28.7  
Max 272.5 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.0 F 0.0  
5:00 216.4 F 0.0  
10:00 231.1 F 0.0  
15:00 230.0 F 0.0  
20:00 229.0 F 0.0  
25:00 228.3 F 0.0  
30:00 225.4 F 0.0  
Drying Complete

FILLING: 1:03  
HEATING: 21:34  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:59:49  
CYCLE COMPLETE

11/10/2022 02:11 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_