
Midmark M11 - v1.0.0
Total Cycles: 4839

Name: RITTER 1
Sterilizer ID: V2178198
08/21/2024 12:08 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.7 F 0.0
2:00 167.1 F 0.9
4:00 187.4 F 2.8
6:00 206.1 F 8.0
8:00 220.6 F 13.0
10:00 227.1 F 7.7
12:00 239.7 F 13.0
14:00 239.3 F 10.6
16:00 236.6 F 8.9
18:00 251.7 F 16.0
20:00 263.8 F 23.1
22:00 271.5 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.1
1:00 272.7 F 28.7
2:00 271.5 F 28.1
3:00 272.7 F 28.8
4:00 271.7 F 28.1

Min 271.4 F 28.0
Max 272.8 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.4 F 0.0
5:00 211.1 F 0.0
10:00 218.9 F 0.0
15:00 221.4 F 0.0
20:00 225.5 F 0.0
25:00 231.2 F 0.0
30:00 228.0 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 22:00
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 01:00:16
CYCLE COMPLETE

08/21/2024 01:08 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____