
Midmark M11 - v1.0.5
Total Cycles: 1843

Name: RITTER2
Sterilizer ID: V2313228
12/15/2022 07:45 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 78.8 F 0.0
2:00 85.9 F 0.0
4:00 81.8 F 0.0
6:00 100.7 F 0.0
8:00 148.4 F 0.2
10:00 176.8 F 0.9
12:00 199.8 F 4.8
14:00 220.9 F 9.1
16:00 231.8 F 13.3
18:00 231.2 F 8.7
20:00 242.8 F 12.0
22:00 241.6 F 11.8
24:00 239.8 F 10.6
26:00 253.4 F 17.3
28:00 264.8 F 24.2

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.7
1:00 272.7 F 29.4
2:00 271.4 F 28.6
3:00 272.5 F 29.6
4:00 272.1 F 29.0

Min 271.1 F 28.5
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 174.4 F 0.0
5:00 211.4 F 0.0
10:00 227.0 F 0.0
15:00 228.8 F 0.0
20:00 221.6 F 0.0
25:00 219.9 F 0.0
30:00 210.9 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 29:53
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 01:07:59
CYCLE COMPLETE

12/15/2022 08:52 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____