

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3641

Name: RITTER2  
Sterilizer ID: V2313228  
05/24/2024 08:09 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.3 F 0.0  
2:00 76.5 F 0.0  
4:00 79.6 F 0.3  
6:00 119.8 F 2.2  
8:00 152.8 F 5.0  
10:00 172.3 F 8.5  
12:00 205.1 F 12.3  
14:00 213.9 F 4.4  
16:00 225.4 F 7.7  
18:00 234.5 F 11.5  
20:00 228.7 F 6.9  
22:00 240.4 F 12.0  
24:00 237.8 F 9.8  
26:00 250.2 F 15.8  
28:00 260.9 F 21.8  
30:00 270.8 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 272.2 F 28.9  
1:00 272.4 F 29.7  
2:00 271.7 F 28.8  
3:00 272.7 F 29.6  
4:00 271.5 F 28.9  
  
Min 271.4 F 28.8  
Max 272.7 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.5 F 0.0  
5:00 217.0 F 0.0  
10:00 191.9 F 0.0  
15:00 189.6 F 0.0  
20:00 186.3 F 0.0  
25:00 183.4 F 0.0  
30:00 181.6 F 0.0  
Drying Complete

FILLING: 0:49  
HEATING: 30:36  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 01:08:38  
CYCLE COMPLETE

05/24/2024 09:17 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_