
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001036
13:52 13/04/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:16
PRESSURIZING 3:20
STERILIZING 4:29
136.4 C 220kPa 4:29
136.4 C 223kPa 4:59
136.3 C 223kPa 5:29
136.3 C 223kPa 5:59
136.2 C 223kPa 6:29
136.2 C 223kPa 6:59
136.4 C 225kPa 7:29
136.4 C 224kPa 7:59
136.4 C 224kPa 7:59
VENTING 7:59
AIR DRYING 8:51
CYCLE COMPLETE 17:06

Digital Signature #
E8D2F7D0B3D3D160
