
Midmark M11 - v1.0.5
Total Cycles: 10

Name: RITTER2
Sterilizer ID: V2313228
04/22/2021 11:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 129.7 F 0.0
2:00 129.7 F 1.6
4:00 146.3 F 4.4
6:00 174.5 F 8.4
8:00 196.6 F 13.0
10:00 212.0 F 6.5
12:00 226.1 F 11.2
14:00 229.6 F 8.5
16:00 239.1 F 11.7
18:00 244.6 F 13.8
20:00 256.7 F 20.4
22:00 268.0 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.5
1:00 271.5 F 29.3
2:00 271.7 F 29.5
3:00 271.9 F 29.7
4:00 271.8 F 29.9

Min 271.3 F 29.3
Max 272.5 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.4 F 0.0
5:00 220.3 F 0.0
10:00 221.2 F 0.0
15:00 210.6 F 0.0
20:00 214.3 F 0.0
25:00 215.7 F 0.0
30:00 221.3 F 0.0
Drying Complete

FILLING: 1:26
HEATING: 23:07
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:01:50
CYCLE COMPLETE

04/22/2021 12:45 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____