

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4343

Name: RITTER2  
Sterilizer ID: V2313228  
02/10/2025 10:22 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 68.3 F 0.0  
2:00 80.1 F 0.0  
4:00 85.5 F 0.3  
6:00 108.7 F 2.2  
8:00 150.8 F 5.6  
10:00 170.5 F 9.3  
12:00 191.1 F 13.1  
14:00 210.4 F 4.1  
16:00 218.6 F 7.5  
18:00 230.0 F 11.3  
20:00 228.8 F 6.8  
22:00 241.6 F 12.1  
24:00 238.6 F 10.1  
26:00 251.3 F 16.3  
28:00 262.3 F 22.6  
30:00 272.7 F 29.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.8  
1:00 272.1 F 28.7  
2:00 272.4 F 28.9  
3:00 272.8 F 29.2  
4:00 272.5 F 29.5

Min 271.7 F 28.7  
Max 272.9 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 185.9 F 0.0  
5:00 202.0 F 0.0  
10:00 198.3 F 0.0  
15:00 205.5 F 0.0  
20:00 221.5 F 0.0  
25:00 223.1 F 0.0  
30:00 220.9 F 0.0  
Drying Complete

FILLING: 0:50  
HEATING: 30:20  
STERILIZING: 4:00  
VENTING: 3:26  
DRYING: 30:00  
TOTAL CYCLE: 01:08:36  
CYCLE COMPLETE

02/10/2025 11:30 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_