
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.6uS / 0.3ppm
CYCLE NUMBER 000512
11:38 21/05/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 0:46
PRESSURIZING 0:57
STERILIZING 2:12
136.4 C 233kPa 2:12
136.3 C 232kPa 2:42
136.4 C 231kPa 3:12
136.2 C 231kPa 3:42
136.3 C 232kPa 4:12
136.3 C 232kPa 4:42
136.4 C 231kPa 5:12
136.4 C 231kPa 5:42
136.4 C 231kPa 5:42
VENTING 5:42
AIR DRYING 6:38
CYCLE COMPLETE 11:28

Digital Signature #
5D7FDAEBD6D494D6
