

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3468

Name: RITTER 1  
Sterilizer ID: V2178198  
06/29/2023 02:00 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 198.2 F 0.0  
2:00 190.6 F 4.2  
4:00 201.3 F 9.3  
6:00 220.4 F 8.5  
8:00 229.4 F 9.9  
10:00 234.3 F 8.1  
12:00 244.4 F 13.5  
14:00 244.2 F 12.4  
16:00 257.2 F 19.3  
18:00 268.7 F 26.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.4  
1:00 272.7 F 29.3  
2:00 272.2 F 28.7  
3:00 271.5 F 28.4  
4:00 272.7 F 29.3  
  
Min 271.5 F 28.3  
Max 272.9 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.8 F 0.0  
5:00 221.4 F 0.0  
10:00 241.8 F 0.0  
15:00 236.1 F 0.0  
20:00 225.2 F 0.0  
25:00 224.4 F 0.0  
30:00 222.2 F 0.0  
Drying Complete

FILLING: 1:50  
HEATING: 19:11  
STERILIZING: 4:00  
VENTING: 3:17  
DRYING: 30:00  
TOTAL CYCLE: 00:58:18  
CYCLE COMPLETE

06/29/2023 02:58 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_