
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001630
11:12 29/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:13
PRESSURIZING 2:18
STERILIZING 3:26
136.0 C 221kPa 3:26
135.9 C 222kPa 3:56
136.3 C 224kPa 4:26
136.3 C 225kPa 4:56
136.4 C 225kPa 5:26
136.4 C 226kPa 5:56
136.4 C 225kPa 6:26
136.4 C 224kPa 6:56
136.4 C 224kPa 6:56
VENTING 6:56
AIR DRYING 7:48
CYCLE COMPLETE 7:52

Digital Signature #
61C4E9FB2612D199
