

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3055

Name: RITTER 1  
Sterilizer ID: V2178198  
03/30/2023 07:24 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 78.9 F 0.0  
2:00 85.5 F 0.0  
4:00 113.2 F 0.0  
6:00 140.8 F 0.0  
8:00 155.5 F 0.3  
10:00 184.2 F 2.3  
12:00 200.1 F 5.6  
14:00 220.1 F 8.8  
16:00 231.8 F 12.4  
18:00 228.6 F 6.9  
20:00 238.4 F 11.5  
22:00 235.6 F 8.4  
24:00 242.7 F 11.0  
26:00 245.5 F 12.8  
28:00 257.5 F 19.3  
30:00 268.3 F 26.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.2  
1:00 272.6 F 28.9  
2:00 271.3 F 28.1  
3:00 272.5 F 29.0  
4:00 272.1 F 28.4  
  
Min 271.3 F 28.1  
Max 272.6 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.7 F 0.0  
5:00 206.5 F 0.0  
10:00 201.9 F 0.0  
15:00 204.8 F 0.0  
20:00 228.0 F 0.0  
25:00 227.1 F 0.0  
30:00 215.2 F 0.0  
Drying Complete

FILLING: 0:46  
HEATING: 31:17  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 01:09:18  
CYCLE COMPLETE

03/30/2023 08:33 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_