
Midmark M11 - v1.0.0
Total Cycles: 2380

Name: RITTER 1
Sterilizer ID: V2178198
09/27/2022 03:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 139.7 F 0.0
2:00 146.7 F 2.4
4:00 170.0 F 6.5
6:00 195.8 F 11.8
8:00 216.2 F 5.5
10:00 230.0 F 10.9
12:00 230.1 F 7.2
14:00 244.4 F 13.6
16:00 243.0 F 12.1
18:00 256.3 F 19.0
20:00 268.3 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.7
1:00 272.7 F 29.4
2:00 271.6 F 28.6
3:00 272.2 F 29.4
4:00 272.5 F 29.2

Min 271.3 F 28.5
Max 272.9 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.6 F 0.0
5:00 207.9 F 0.0
10:00 206.3 F 0.0
15:00 211.4 F 0.0
20:00 225.8 F 0.0
25:00 234.3 F 0.0
30:00 226.8 F 0.0
Drying Complete

FILLING: 1:44
HEATING: 21:13
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:00:03
CYCLE COMPLETE

09/27/2022 04:00 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____