
Midmark M11 - v1.0.0
Total Cycles: 2024

Name: RITTER 1
Sterilizer ID: V2178198
06/15/2022 09:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.5 F 0.0
2:00 184.8 F 3.1
4:00 184.1 F 6.8
6:00 210.9 F 12.1
8:00 222.8 F 6.2
10:00 234.5 F 11.3
12:00 233.7 F 7.9
14:00 243.8 F 12.0
16:00 244.2 F 12.3
18:00 256.8 F 18.9
20:00 267.7 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.2 F 28.7
2:00 272.0 F 28.9
3:00 271.7 F 28.4
4:00 272.7 F 29.1

Min 271.3 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.4 F 0.0
5:00 227.1 F 0.0
10:00 238.5 F 0.0
15:00 227.7 F 0.0
20:00 225.7 F 0.0
25:00 225.4 F 0.0
30:00 225.4 F 0.0
Drying Complete

FILLING: 1:00
HEATING: 21:19
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:59:30
CYCLE COMPLETE

06/15/2022 10:51 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____