
Midmark M11 - v1.0.5
Total Cycles: 1887

Name: RITTER2
Sterilizer ID: V2313228
12/28/2022 01:57 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 132.8 F 0.0
2:00 134.3 F 0.0
4:00 165.3 F 0.0
6:00 193.0 F 4.3
8:00 210.9 F 9.1
10:00 222.0 F 9.6
12:00 233.6 F 10.0
14:00 235.8 F 8.8
16:00 243.3 F 11.7
18:00 247.3 F 13.9
20:00 260.3 F 21.2
22:00 272.0 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.7 F 29.1
2:00 271.5 F 28.5
3:00 272.9 F 29.2
4:00 271.7 F 28.4

Min 271.5 F 28.3
Max 272.9 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 171.9 F 0.0
5:00 208.3 F 0.0
10:00 227.9 F 0.0
15:00 223.2 F 0.0
20:00 218.9 F 0.0
25:00 215.6 F 0.0
30:00 211.1 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 22:31
STERILIZING: 4:00
VENTING: 3:00
DRYING: 30:00
TOTAL CYCLE: 01:00:26
CYCLE COMPLETE

12/28/2022 02:57 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____