

-----

Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 006160  
10:21 18/08/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:56  
PRESSURIZING 1:59  
STERILIZING 3:04  
136.0 C 220kPa 3:04  
136.4 C 223kPa 3:34  
136.4 C 221kPa 4:04  
136.1 C 222kPa 4:34  
136.4 C 221kPa 5:04  
136.4 C 223kPa 5:34  
136.3 C 224kPa 6:04  
136.3 C 220kPa 6:34  
136.3 C 220kPa 6:34  
VENTING 6:34  
AIR DRYING 7:24  
CYCLE COMPLETE 17:18

-----

Digital Signature #  
B67B503945400110  
-----