

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4122

Name: RITTER 1  
Sterilizer ID: V2178198  
12/15/2023 08:42 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 71.6 F 0.0  
2:00 76.7 F 0.2  
4:00 83.0 F 1.0  
6:00 107.9 F 3.1  
8:00 138.8 F 5.8  
10:00 165.8 F 9.1  
12:00 186.3 F 12.9  
14:00 211.8 F 3.7  
16:00 221.6 F 6.9  
18:00 228.3 F 10.5  
20:00 231.2 F 8.1  
22:00 238.4 F 10.3  
24:00 236.7 F 8.6  
26:00 246.6 F 13.3  
28:00 257.4 F 19.1  
30:00 267.6 F 25.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.1  
1:00 272.5 F 28.7  
2:00 271.2 F 28.1  
3:00 272.4 F 28.8  
4:00 271.4 F 28.0

Min 271.1 F 27.9  
Max 272.6 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 182.3 F 0.0  
5:00 242.0 F 0.0  
10:00 221.4 F 0.0  
15:00 224.9 F 0.0  
20:00 229.8 F 0.0  
25:00 222.7 F 0.0  
30:00 216.1 F 0.0  
Drying Complete

FILLING: 0:53  
HEATING: 31:26  
STERILIZING: 4:00  
VENTING: 3:23  
DRYING: 30:00  
TOTAL CYCLE: 01:09:42  
CYCLE COMPLETE

12/15/2023 09:51 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_