
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.8uS / 0.5ppm
CYCLE NUMBER 001671
10:46 01/06/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:01
PRESSURIZING 3:04
STERILIZING 4:14
136.4 C 222kPa 4:14
136.3 C 225kPa 4:44
136.4 C 225kPa 5:14
136.4 C 224kPa 5:44
136.4 C 224kPa 6:14
136.2 C 221kPa 6:44
136.4 C 224kPa 7:14
136.1 C 222kPa 7:44
136.1 C 222kPa 7:44
VENTING 7:45
AIR DRYING 8:39
CYCLE COMPLETE 8:51

Digital Signature #
2FC79370601D9128
