
Midmark M11 - v1.0.0
Total Cycles: 791

Name: RITTER 1
Sterilizer ID: V2178198
04/21/2021 09:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 140.3 F 0.0
2:00 139.1 F 1.0
4:00 147.0 F 3.8
6:00 180.2 F 7.9
8:00 205.7 F 13.2
10:00 231.8 F 8.1
12:00 243.0 F 13.5
14:00 239.9 F 10.4
16:00 234.6 F 7.8
18:00 249.3 F 14.7
20:00 261.2 F 21.4
22:00 272.1 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.7
1:00 271.9 F 29.1
2:00 272.4 F 29.2
3:00 272.6 F 29.1
4:00 272.6 F 29.1

Min 271.8 F 28.6
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.5 F 0.0
5:00 240.4 F 0.0
10:00 228.2 F 0.0
15:00 226.3 F 0.0
20:00 226.0 F 0.0
25:00 226.8 F 0.0
30:00 224.2 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 23:05
STERILIZING: 4:00
VENTING: 2:59
DRYING: 30:00
TOTAL CYCLE: 01:01:01
CYCLE COMPLETE

04/21/2021 10:40 AM
ACCEPTED: X
REJECTED:
Operator: U01