

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3874

Name: RITTER 1  
Sterilizer ID: V2178198  
10/05/2023 02:26 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 159.5 F 0.0  
2:00 163.3 F 3.0  
4:00 195.4 F 6.8  
6:00 212.2 F 11.7  
8:00 226.7 F 6.5  
10:00 233.3 F 11.5  
12:00 234.5 F 8.3  
14:00 242.2 F 11.5  
16:00 245.3 F 12.8  
18:00 257.3 F 19.2  
20:00 267.8 F 25.9

STERILIZING  
mm:ss Degrees PSI  
0:00 272.5 F 28.6  
1:00 272.3 F 28.7  
2:00 272.0 F 28.7  
3:00 272.1 F 28.8  
4:00 272.3 F 29.0

Min 271.6 F 28.5  
Max 272.8 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 177.8 F 0.0  
5:00 192.6 F 0.0  
10:00 194.0 F 0.0  
15:00 194.8 F 0.0  
20:00 193.3 F 0.0  
25:00 190.9 F 0.0  
30:00 189.8 F 0.0  
Drying Complete

FILLING: 1:58  
HEATING: 21:42  
STERILIZING: 4:00  
VENTING: 3:43  
DRYING: 30:00  
TOTAL CYCLE: 01:01:23  
CYCLE COMPLETE

10/05/2023 03:27 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_