

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 006963  
7:55 22/02/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:29  
PRESSURIZING 1:34  
STERILIZING 2:37  
136.4 C 221kPa 2:37  
136.4 C 223kPa 3:07  
136.4 C 222kPa 3:37  
136.4 C 220kPa 4:07  
136.5 C 222kPa 4:37  
136.1 C 220kPa 5:07  
136.4 C 221kPa 5:37  
136.2 C 221kPa 6:07  
136.2 C 221kPa 6:07  
VENTING 6:07  
AIR DRYING 6:56  
CYCLE COMPLETE 7:03

-----  
Digital Signature #  
F450CAD355706CDF  
-----