
Midmark M11 - v1.0.0
Total Cycles: 5053

Name: RITTER 1
Sterilizer ID: V2178198
11/04/2024 03:44 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 153.5 F 0.0
2:00 152.9 F 0.2
4:00 172.1 F 0.7
6:00 190.8 F 3.6
8:00 207.9 F 7.6
10:00 221.1 F 11.8
12:00 226.6 F 6.8
14:00 237.2 F 11.9
16:00 237.0 F 9.5
18:00 238.7 F 9.4
20:00 248.7 F 14.4
22:00 260.3 F 20.8
24:00 270.8 F 27.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.0
1:00 272.5 F 28.8
2:00 271.7 F 28.2
3:00 271.6 F 28.4
4:00 272.4 F 28.7

Min 271.2 F 27.9
Max 272.5 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.2 F 0.0
5:00 194.7 F 0.0
10:00 192.4 F 0.0
15:00 191.9 F 0.0
20:00 198.1 F 0.0
25:00 207.7 F 0.0
30:00 211.6 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 24:48
STERILIZING: 4:00
VENTING: 3:48
DRYING: 30:00
TOTAL CYCLE: 01:03:33
CYCLE COMPLETE

11/04/2024 04:47 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____