
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 007364
8:09 13/05/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:53
PRESSURIZING 2:59
STERILIZING 4:01
136.0 C 220kPa 4:01
136.3 C 221kPa 4:31
136.0 C 222kPa 5:01
136.4 C 224kPa 5:31
136.3 C 223kPa 6:01
136.1 C 223kPa 6:31
135.8 C 221kPa 7:01
136.1 C 221kPa 7:31
136.1 C 221kPa 7:31
VENTING 7:31
AIR DRYING 8:23
CYCLE COMPLETE 8:31

Digital Signature #
62478360E7B1B648
