
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.6uS / 0.3ppm
CYCLE NUMBER 000504
14:09 07/05/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:18
PRESSURIZING 3:23
STERILIZING 4:33
136.0 C 229kPa 4:33
136.2 C 232kPa 5:03
136.4 C 233kPa 5:33
136.2 C 231kPa 6:03
136.4 C 233kPa 6:33
136.3 C 232kPa 7:03
136.4 C 233kPa 7:33
136.4 C 233kPa 8:03
136.4 C 233kPa 8:03
VENTING 8:03
AIR DRYING 8:59
CYCLE COMPLETE 11:26

Digital Signature #
CCBAED8D853C2549
