

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000708  
9:03 20/09/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:58  
PRESSURIZING 3:03  
STERILIZING 4:08  
136.4 C 220kPa 4:08  
136.2 C 231kPa 4:38  
136.2 C 231kPa 5:08  
136.4 C 233kPa 5:38  
136.4 C 233kPa 6:08  
136.4 C 233kPa 6:38  
136.3 C 232kPa 7:08  
136.4 C 233kPa 7:38  
136.4 C 233kPa 7:38  
VENTING 7:38  
AIR DRYING 8:33  
CYCLE COMPLETE 8:36

-----  
Digital Signature #  
087CC179C1C07F6E  
-----