
Midmark M11 - v1.0.5
Total Cycles: 1732

Name: RITTER2
Sterilizer ID: V2313228
11/16/2022 11:19 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 155.5 F 0.0
2:00 156.4 F 0.2
4:00 181.6 F 0.8
6:00 206.7 F 5.4
8:00 219.3 F 10.7
10:00 225.6 F 6.6
12:00 239.0 F 12.7
14:00 239.3 F 11.1
16:00 238.8 F 10.2
18:00 253.1 F 17.3
20:00 265.1 F 24.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.9
1:00 272.6 F 29.5
2:00 271.5 F 29.0
3:00 272.6 F 29.6
4:00 271.7 F 28.9

Min 271.5 F 28.8
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.8 F 0.0
5:00 215.6 F 0.0
10:00 225.5 F 0.0
15:00 230.2 F 0.0
20:00 228.7 F 0.0
25:00 229.7 F 0.0
30:00 228.6 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 21:46
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:59:58
CYCLE COMPLETE

11/16/2022 12:18 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____