

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002004  
15:01 12/07/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:38  
PRESSURIZING 3:39  
STERILIZING 4:30  
136.0 C 220kPa 4:30  
136.2 C 224kPa 5:00  
136.1 C 224kPa 5:30  
136.1 C 219kPa 6:00  
136.3 C 223kPa 6:30  
136.4 C 225kPa 7:00  
136.4 C 224kPa 7:30  
136.2 C 222kPa 8:00  
136.2 C 222kPa 8:00  
VENTING 8:01  
AIR DRYING 9:00  
CYCLE COMPLETE 9:01

-----  
Digital Signature #  
F1A4EB716A6DE769  
-----