
Midmark M11 - v1.0.0
Total Cycles: 2319

Name: RITTER 1
Sterilizer ID: V2178198
09/13/2022 12:09 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 157.1 F 0.0
2:00 155.7 F 2.6
4:00 170.3 F 7.1
6:00 195.9 F 12.4
8:00 218.8 F 6.3
10:00 236.1 F 11.4
12:00 230.8 F 7.7
14:00 244.6 F 12.9
16:00 242.7 F 12.1
18:00 255.9 F 18.8
20:00 267.4 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.6
1:00 272.0 F 28.6
2:00 272.5 F 28.9
3:00 272.6 F 29.2
4:00 271.8 F 28.7

Min 271.5 F 28.5
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.6 F 0.0
5:00 229.3 F 0.0
10:00 241.9 F 0.0
15:00 230.5 F 0.0
20:00 219.5 F 0.0
25:00 232.0 F 0.0
30:00 213.8 F 0.0
Drying Complete

FILLING: 1:37
HEATING: 21:14
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:00:06
CYCLE COMPLETE

09/13/2022 01:09 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____