
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.7uS / 0.4ppm
CYCLE NUMBER 005073
15:03 06/02/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:52
PRESSURIZING 2:53
STERILIZING 3:49
136.4 C 224kPa 3:49
136.4 C 224kPa 4:19
136.2 C 224kPa 4:49
136.4 C 223kPa 5:19
136.1 C 218kPa 5:49
136.4 C 224kPa 6:19
136.4 C 225kPa 6:49
136.1 C 219kPa 7:19
136.1 C 219kPa 7:19
VENTING 7:19
AIR DRYING 8:09
CYCLE COMPLETE 8:13

Digital Signature #
FE0943F121D11465
