

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 001941  
8:16 12/08/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:52  
PRESSURIZING 3:54  
STERILIZING 5:02  
136.0 C 221kPa 5:02  
136.3 C 220kPa 5:32  
136.4 C 221kPa 6:02  
136.4 C 224kPa 6:32  
136.1 C 221kPa 7:02  
136.4 C 223kPa 7:32  
136.4 C 224kPa 8:02  
136.1 C 222kPa 8:32  
136.1 C 222kPa 8:32  
VENTING 8:32  
AIR DRYING 9:24  
CYCLE COMPLETE 9:37  
-----

Digital Signature #  
AD9A8E1FD63E521E  
-----