
Midmark M11 - v1.0.0
Total Cycles: 5504

Name:
Sterilizer ID: V2178198
04/14/2025 11:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 90.9 F 0.0
2:00 91.9 F 0.1
4:00 106.1 F 2.0
6:00 141.4 F 4.5
8:00 164.3 F 7.1
10:00 182.6 F 10.2
12:00 216.7 F 6.5
14:00 210.7 F 5.1
16:00 221.4 F 8.7
18:00 232.8 F 12.8
20:00 234.9 F 8.7
22:00 245.0 F 12.4
24:00 243.2 F 11.9
26:00 254.4 F 17.8
28:00 264.9 F 24.0
30:00 270.9 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.8
1:00 272.4 F 29.6
2:00 272.8 F 29.3
3:00 271.2 F 28.7
4:00 271.8 F 29.3

Min 271.1 F 28.7
Max 273.2 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.7 F 0.0
5:00 209.9 F 0.0
10:00 219.8 F 0.0
15:00 245.6 F 0.0
20:00 234.8 F 0.0
25:00 225.1 F 0.0
30:00 229.0 F 0.0
Drying Complete

FILLING: 0:44
HEATING: 30:51
STERILIZING: 4:00
VENTING: 3:58
DRYING: 30:00
TOTAL CYCLE: 01:09:34
CYCLE COMPLETE

04/14/2025 01:08 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____