

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3678

Name: RITTER2  
Sterilizer ID: V2313228  
06/07/2024 01:32 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 117.7 F 0.0  
2:00 120.2 F 1.0  
4:00 130.4 F 3.2  
6:00 167.3 F 6.5  
8:00 186.0 F 10.3  
10:00 225.4 F 8.3  
12:00 221.6 F 7.0  
14:00 235.8 F 11.0  
16:00 231.8 F 7.5  
18:00 241.7 F 12.1  
20:00 238.5 F 10.2  
22:00 250.1 F 16.0  
24:00 260.8 F 22.0  
26:00 270.6 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.6  
1:00 271.3 F 29.5  
2:00 271.5 F 29.5  
3:00 271.8 F 29.6  
4:00 272.7 F 29.8

Min 271.1 F 29.5  
Max 272.9 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 180.5 F 0.0  
5:00 230.4 F 0.0  
10:00 217.4 F 0.0  
15:00 217.5 F 0.0  
20:00 173.1 F 0.0  
25:00 171.9 F 0.0  
30:00 175.9 F 0.0  
Drying Complete

FILLING: 0:57  
HEATING: 27:32  
STERILIZING: 4:00  
VENTING: 3:42  
DRYING: 30:00  
TOTAL CYCLE: 01:06:11  
CYCLE COMPLETE

06/07/2024 02:38 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_