

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 005324  
14:19 04/04/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:29  
PRESSURIZING 2:32  
STERILIZING 3:37  
136.4 C 221kPa 3:37  
136.4 C 225kPa 4:07  
136.2 C 224kPa 4:37  
136.3 C 225kPa 5:07  
136.3 C 223kPa 5:37  
136.2 C 223kPa 6:07  
136.4 C 225kPa 6:37  
136.4 C 225kPa 7:07  
136.4 C 225kPa 7:07  
VENTING 7:07  
AIR DRYING 8:03  
CYCLE COMPLETE 8:14

-----  
Digital Signature #  
029570D26C10B78C  
-----