
Midmark M11 - v1.0.5
Total Cycles: 2765

Name: RITTER2
Sterilizer ID: V2313228
08/14/2023 07:13 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 181.3 F 0.0
2:00 179.7 F 0.4
4:00 201.2 F 5.9
6:00 219.5 F 11.2
8:00 226.2 F 7.0
10:00 238.6 F 12.7
12:00 239.6 F 10.9
14:00 238.7 F 10.1
16:00 253.3 F 17.3
18:00 265.6 F 24.6

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.7
1:00 272.0 F 28.8
2:00 272.6 F 29.2
3:00 272.4 F 29.5
4:00 271.6 F 28.5

Min 271.5 F 28.5
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.6 F 0.0
5:00 223.1 F 0.0
10:00 226.1 F 0.0
15:00 228.2 F 0.0
20:00 225.1 F 0.0
25:00 224.2 F 0.0
30:00 222.0 F 0.0

Drying Complete

FILLING: 2:10
HEATING: 19:35
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:54
CYCLE COMPLETE

08/14/2023 08:11 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____