
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 002236
8:42 11/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:26
PRESSURIZING 2:35
STERILIZING 3:39
136.4 C 220kPa 3:39
136.0 C 223kPa 4:09
136.4 C 224kPa 4:39
136.4 C 224kPa 5:09
136.2 C 221kPa 5:39
136.4 C 223kPa 6:09
136.1 C 221kPa 6:39
136.2 C 220kPa 7:09
136.2 C 220kPa 7:09
VENTING 7:10
AIR DRYING 8:01
CYCLE COMPLETE 29:31

Digital Signature #
02C0DE08A6F3505B
