

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2953

Name: RITTER2  
Sterilizer ID: V2313228  
09/29/2023 02:51 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 179.2 F 0.0  
2:00 171.1 F 0.3  
4:00 191.4 F 5.0  
6:00 211.5 F 10.1  
8:00 226.2 F 6.5  
10:00 236.1 F 11.3  
12:00 235.4 F 8.9  
14:00 239.3 F 9.9  
16:00 248.4 F 14.6  
18:00 260.5 F 21.4  
20:00 271.5 F 28.6

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 28.6  
1:00 272.6 F 29.2  
2:00 271.4 F 29.0  
3:00 271.9 F 28.8  
4:00 272.3 F 29.4

Min 271.2 F 28.5  
Max 272.7 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 178.1 F 0.0  
5:00 245.4 F 0.0  
10:00 223.2 F 0.0  
15:00 233.0 F 0.0  
20:00 226.5 F 0.0  
25:00 230.8 F 0.0  
30:00 229.5 F 0.0  
Drying Complete

FILLING: 0:57  
HEATING: 20:28  
STERILIZING: 4:00  
VENTING: 3:21  
DRYING: 30:00  
TOTAL CYCLE: 00:58:46  
CYCLE COMPLETE

09/29/2023 03:49 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_