
Midmark M11 - v1.0.5
Total Cycles: 2192

Name: RITTER2
Sterilizer ID: V2313228
03/29/2023 10:21 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 161.0 F 0.0
2:00 158.6 F 3.1
4:00 173.3 F 6.8
6:00 196.0 F 11.5
8:00 214.7 F 5.2
10:00 225.6 F 10.3
12:00 231.3 F 7.7
14:00 242.2 F 13.2
16:00 242.5 F 12.1
18:00 255.6 F 19.0
20:00 267.0 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.3
1:00 272.5 F 30.0
2:00 271.7 F 29.2
3:00 272.0 F 30.0
4:00 272.3 F 29.8

Min 271.3 F 29.1
Max 272.6 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.1 F 0.0
5:00 215.1 F 0.0
10:00 223.2 F 0.0
15:00 222.3 F 0.0
20:00 227.8 F 0.0
25:00 229.3 F 0.0
30:00 228.5 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 21:28
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:59:47
CYCLE COMPLETE

03/29/2023 11:20 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____