
Midmark M11 - v1.0.0
Total Cycles: 4368

Name: RITTER 1
Sterilizer ID: V2178198
03/07/2024 05:34 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 160.6 F 0.0
2:00 163.4 F 3.1
4:00 181.0 F 7.0
6:00 203.5 F 11.5
8:00 217.0 F 5.4
10:00 228.1 F 10.4
12:00 231.8 F 7.6
14:00 242.8 F 12.7
16:00 241.4 F 11.2
18:00 254.1 F 17.5
20:00 264.9 F 24.0

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.4
1:00 272.5 F 29.1
2:00 271.4 F 28.3
3:00 272.6 F 29.0
4:00 271.4 F 28.3

Min 271.3 F 28.3
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.1 F 0.0
5:00 229.4 F 0.0
10:00 236.2 F 0.0
15:00 227.0 F 0.0
20:00 223.9 F 0.0
25:00 226.7 F 0.0
30:00 223.5 F 0.0
Drying Complete

FILLING: 1:47
HEATING: 21:45
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:01:07
CYCLE COMPLETE

03/07/2024 06:35 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____