
Midmark M11 - v1.0.0
Total Cycles: 1004

Name: RITTER 1
Sterilizer ID: V2178198
07/13/2021 01:10 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 198.8 F 0.0
2:00 194.2 F 3.3
4:00 197.3 F 7.6
6:00 216.5 F 13.1
8:00 226.2 F 8.0
10:00 239.7 F 13.3
12:00 239.6 F 10.9
14:00 237.9 F 9.3
16:00 251.7 F 16.2
18:00 263.4 F 23.1
20:00 271.9 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.4
1:00 272.8 F 29.0
2:00 271.7 F 28.4
3:00 272.8 F 29.1
4:00 272.1 F 28.4

Min 271.6 F 28.3
Max 273.0 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.2 F 0.0
5:00 227.2 F 0.0
10:00 234.6 F 0.0
15:00 231.8 F 0.0
20:00 227.7 F 0.0
25:00 228.2 F 0.0
30:00 213.3 F 0.0
Drying Complete

FILLING: 1:35
HEATING: 20:06
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:51
CYCLE COMPLETE

07/13/2021 02:09 PM
ACCEPTED: X
REJECTED:
Operator: U01