

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3632

Name: RITTER2  
Sterilizer ID: V2313228  
05/21/2024 03:05 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 82.4 F 0.0  
2:00 83.2 F 0.0  
4:00 92.8 F 0.8  
6:00 120.9 F 3.0  
8:00 148.0 F 5.8  
10:00 170.7 F 8.8  
12:00 187.5 F 12.0  
14:00 205.2 F 4.3  
16:00 218.2 F 7.7  
18:00 228.5 F 11.5  
20:00 228.7 F 7.7  
22:00 240.2 F 12.8  
24:00 239.6 F 10.9  
26:00 250.5 F 16.3  
28:00 260.0 F 22.0  
30:00 269.6 F 28.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 30.0  
1:00 272.0 F 29.9  
2:00 272.3 F 30.0  
3:00 272.3 F 30.3  
4:00 272.1 F 30.4  
  
Min 271.5 F 29.9  
Max 272.8 F 30.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.4 F 0.0  
5:00 215.9 F 0.0  
10:00 223.0 F 0.0  
15:00 223.5 F 0.0  
20:00 227.1 F 0.0  
25:00 229.1 F 0.0  
30:00 226.8 F 0.0  
Drying Complete

FILLING: 0:54  
HEATING: 31:55  
STERILIZING: 4:00  
VENTING: 3:37  
DRYING: 30:00  
TOTAL CYCLE: 01:10:27  
CYCLE COMPLETE

05/21/2024 04:15 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_