
Midmark M11 - v1.0.5
Total Cycles: 124

Name: RITTER2
Sterilizer ID: V2313228
06/22/2021 06:49 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.5 F 0.0
2:00 72.9 F 0.2
4:00 77.0 F 1.1
6:00 96.6 F 3.2
8:00 127.9 F 6.2
10:00 161.7 F 9.8
12:00 183.6 F 13.4
14:00 201.4 F 5.3
16:00 214.1 F 9.1
18:00 226.4 F 13.1
20:00 229.1 F 9.6
22:00 237.3 F 10.3
24:00 241.1 F 13.1
26:00 253.0 F 18.9
28:00 263.6 F 25.1
30:00 272.7 F 31.1

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 30.5
1:00 273.3 F 31.3
2:00 272.7 F 30.6
3:00 272.3 F 30.7
4:00 273.1 F 31.3

Min 271.7 F 30.3
Max 273.4 F 31.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.0 F 0.0
5:00 229.5 F 0.0
10:00 224.1 F 0.0
15:00 209.6 F 0.0
20:00 202.6 F 0.0
25:00 199.0 F 0.0
30:00 199.9 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 30:20
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:08:34
CYCLE COMPLETE

06/22/2021 07:57 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____