
Midmark M11 - v1.0.5
Total Cycles: 825

Name: RITTER2
Sterilizer ID: V2313228
02/23/2022 10:16 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 188.1 F 0.0
2:00 185.5 F 3.8
4:00 200.5 F 9.4
6:00 221.2 F 7.1
8:00 232.7 F 11.3
10:00 235.2 F 9.6
12:00 237.0 F 9.1
14:00 250.6 F 16.1
16:00 262.9 F 23.3
18:00 272.0 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.1
1:00 272.7 F 29.6
2:00 271.7 F 29.3
3:00 272.7 F 29.6
4:00 271.6 F 29.0

Min 271.5 F 29.0
Max 273.0 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.1 F 0.0
5:00 217.3 F 0.0
10:00 242.9 F 0.0
15:00 228.1 F 0.0
20:00 223.2 F 0.0
25:00 220.8 F 0.0
30:00 219.4 F 0.0
Drying Complete

FILLING: 1:56
HEATING: 18:08
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:57:10
CYCLE COMPLETE

02/23/2022 11:13 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____