
Midmark M11 - v1.0.0
Total Cycles: 3006

Name: RITTER 1
Sterilizer ID: V2178198
03/15/2023 08:45 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.4 F 0.0
2:00 185.7 F 2.3
4:00 188.5 F 6.4
6:00 211.9 F 11.0
8:00 219.1 F 5.5
10:00 232.4 F 11.0
12:00 232.3 F 7.8
14:00 242.6 F 11.5
16:00 245.2 F 13.0
18:00 257.9 F 19.8
20:00 269.2 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.7 F 29.3
2:00 272.2 F 28.7
3:00 271.5 F 28.5
4:00 272.6 F 29.3

Min 271.4 F 28.3
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.0 F 0.0
5:00 219.0 F 0.0
10:00 237.2 F 0.0
15:00 235.6 F 0.0
20:00 226.0 F 0.0
25:00 223.8 F 0.0
30:00 228.7 F 0.0
Drying Complete

FILLING: 1:18
HEATING: 21:05
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:59:36
CYCLE COMPLETE

03/15/2023 09:44 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____