
Midmark M11 - v1.0.5
Total Cycles: 3394

Name: RITTER2
Sterilizer ID: V2313228
02/21/2024 08:48 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	68.3 F	0.0
2:00	84.8 F	0.0
4:00	85.8 F	0.0
6:00	102.4 F	0.4
8:00	144.4 F	0.3
10:00	172.4 F	0.5
12:00	193.6 F	0.9
14:00	213.2 F	1.4
16:00	217.8 F	2.0
18:00	221.2 F	2.8
20:00	221.4 F	2.8
22:00	221.9 F	3.0
24:00	222.1 F	3.0
26:00	222.1 F	3.0
28:00	222.6 F	3.2

C633: HEATUP MODE
PRESSURE LOW
CHAMBER TEMP 224F
CHAMBER PRESS 3.6PSI
** ITEMS NOT STERILE **

02/21/2024 09:18 AM